

Request for Quote

Dr. Nachiketa Tiwari
Department of Mechanical Engineering
IIT Kanpur 208 016.

Tel No:+91-512-6526

Enquiry no.: Acoustics Lab/2015/07

Enquiry Date: 22/12/2015

Closing date: 04/01/2016

Sealed quotations should reach the undersigned latest by 12 noon on January 4, 2016 for array microphones with following specifications.

- ¼ inch free-field microphone
- Pre-polarized with integral pre-amplifier
- BNC or 10-32 connector
- Accuracy: ± 2 dB or better for 50 to 10,000 Hz
- Phase match: ± 5 degrees or better
- Designed for array applications
- TEDS compliant
- Dynamic range (3% distortion at max): 122 dB or better
- Sensitivity: 20 mV/Pa or higher

Proprietary certificate, if relevant, should be provided.

Please provide quotations for 10, 15 and 20 units.

1. Delivery time should be clearly mentioned in the quotation.
2. The quotation should be accompanied with complete product description, technical literature, etc.
3. You are requested to submit your quotation in sealed envelope with complete product description, technical literature, price, warranty period, delivery time and other conditions by 12 noon on January 04, 2016 at the address given below.

Dr. Nachiketa Tiwari
Department of Mechanical Engineering
Indian Institute of Technology
Kanpur 208016.