

Request for Quote

Dr. Nachiketa Tiwari
Department of Mechanical Engineering
IIT Kanpur 208 016.

Tel No:+91-512-6526

Enquiry no.: Acoustics Lab/2016/02

Enquiry Date: 28/04/2016

Closing date: 10/05/2016

Sealed quotations should reach the undersigned latest by 2 PM on May 10, 2016 for array microphones with following specifications.

- ¼ inch free-field microphone
- Pre-polarized with integral pre-amplifier
- BNC or 10-32 connector
- Accuracy: ± 2 dB or better for 50 to 10,000 Hz
- Phase match: ± 5 degrees or better
- Designed for array applications
- TEDS compliant
- Dynamic range (3% distortion at max): 122 dB or better
- Sensitivity: 20 mV/Pa or higher

Proprietary certificate, if relevant, should be provided.

Please provide quotations for 10, 15 and 20 units.

1. Delivery time should be clearly mentioned in the quotation.
2. If microphones are available with BNC as well as 10-32 connector versions, then please quote both versions.
3. The quotation should be accompanied with complete product description, technical literature, etc.
4. You are requested to submit your quotation in sealed envelope with complete product description, technical literature, price, warranty period, delivery time and other conditions by 5 PM on March 12, 2016 at the address given below.

Dr. Nachiketa Tiwari
Department of Mechanical Engineering
Indian Institute of Technology
Kanpur 208016.