

INDIAN INSTITUTE OF TECHNOLOGY KANPUR

Department of Electrical Engineering

Enquiry No.: **EE/SA/INQ/2017-18/16**

Opening Date: 2-Feb-2018

Closing Date: 22-Feb-2018

Sub: Inquiry for Current Probes to be used with Frequency Response Analyzer

We are interested in purchase of current probes with the following configuration to be used with Frequency Response Analyzer. Our organization is an educational institute of repute and liable to get **educational discount** from the manufacturer / supplier. Please specify the discount separately.

Please send your **Sealed Quotation** to the undersigned for the same. The envelope should be marked as "Differential Voltage Probe - **EE/SA/INQ/2017-18/16**"

Items required:

Item required	Specifications	Quantity
Current Probe	Should have standard bnc connector for connecting it to any oscilloscopes, should draw its power for operation from an external power supply / adapter, Power adapter/supply required for this probe should also be included in the quotation , Rating at least 20A _{RMS} continuous for frequency upto 50MHz, at least 100A Peak Pulse, Bandwidth \geq 50MHz, operating temperature range 0°C to +50°C , Should have degauss and dc offset nullification option, Should have at least 3 year warranty .	2

Note:

1. In case you are not manufacturer of the product, your quotation shall contain Authorization Letter from manufacturer.
2. Quotation must be valid for minimum of 60 days.
3. Delivery period should not be more than 8 weeks. And delivery should be CIF New Delhi or CIF IIT Kanpur
4. IIT Kanpur is exempted under 51/96 from payment of IGST. Suitable CDEC certificate can be provided by IIT Kanpur if required.
5. As per Notification No. 45/2017-Central Tax (Rate) dated 14/11/2017, IIT Kanpur has been allowed for GST rate of 5%. Suitable certificate can be provided by IIT Kanpur if required.
6. Send complete detail / brochure of the product(s).
7. Payments terms: Letter of credit (LC) or 100% after successful installation and commissioning.
8. Price must include installation, taxes and all charges.
9. Ambiguous offers, without suitable technical documentation, not mapping compliance to required specifications, may be rejected without any further notification

10. Mere compliance is not sufficient, the technical details must be supported by detailed technical datasheets of the offered product(s)
11. The Institute reserves the right of accepting and rejecting any quotations without assigning any reason.

Sandeep Anand
Department of Electrical Engineering, IIT Kanpur
Kanpur, UP – 208016, India
Email: asandeep@iitk.ac.in