



विद्युत अभियांत्रिकी विभाग
DEPARTMENT OF ELECTRICAL ENGINEERING
भारतीय प्रौद्योगिकी संस्थान कानपुर
INDIAN INSTITUTE OF TECHNOLOGY KANPUR
कानपुर - 208 016 (भारत)
KANPUR - 208 016 (INDIA)

Phone : (0512)-2596523
2597403
Fax : (0512)-2590063
Webpage : <http://www.iitk.ac.in/ee>

ENQ No.: EE/MJA/Osc/2012

Date: January 6, 2012

Inviting sealed quotation for the High Frequency Oscilloscope

Sealed Quotations are invited for the **High Frequency Oscilloscope** with the specifications given below:

Frequency range	Up to 3 GHz
Maximum Allowable Sample Rate (2-Ch)	20 GSa/s
Number of Channels	4
Typical rise time (10% - 90%)	80 ps
Vertical Resolution	8 bits
Typical Channel-to-Channel Isolation	20 dB
Input Impedance	1 M Ω , 50 Ω
Input Sensitivity	1 M Ω : 1 mV/div to 5 V/div 50 Ω : 1 mV/div to 1 V/div
Connectivity	GPIB, LAN, USB

Last date for submission of quote is 16.01.2012.

The Quotation in sealed envelope should be sent to the following address:

Dr. M. J. Akhtar
Asst. Professor
ACES-326
Department of Electrical Engineering
IIT Kanpur, Kanpur-208 016