



Environmental Geochemistry Laboratory
Department of Civil Engineering
Indian Institute of Technology Kanpur

Inquiry no.: IITK/CE/AS/06-2015/01
Closing date and time: 15.06.2015 at 5 pm

Date: 03.06.2015

Sub: Call for quotation for supply and installation of a benchtop pH meter

Sealed quotations (**Technical and Financial bids separately**) are invited from authorized suppliers for items and their specifications given below before 5 pm of 15.06.2015.

The quotation for supply and installation of a **benchtop pH meter** should be sent in two parts in sealed envelopes, clearly marked as "**Technical Bid**" and "**Financial Bid**". The Technical Bid should contain detailed technical specifications of the product being offered and **should not mention any prices**. The Financial Bid should include the detailed price quotation clearly, including the cost of the equipment, taxes, service charges, shipping and handling charges, if any. Our organization is an educational institute of repute and liable to get maximum education discount from manufacturer. Please specify it, separately.

S. No.	Product Name & Specifications	Quantity
1.	<p>Benchtop pH meter with Electrode Stand</p> <p>The pH meter should be a microprocessor-controlled, user friendly system with ease of operation to analyse routine environmental samples from diverse matrices for pH, ORP, Ion, and Temperature.</p> <p>pH:</p> <ul style="list-style-type: none">• Range: 0.0 to 14.0 pH or wider range• Resolution: ± 0.01 pH or better (selectable)• Accuracy: ± 0.01 pH or better• Calibration with 3 points or higher• Temperature compensation: Automatic, 0.0 to 100.0 °C <p>ORP:</p> <ul style="list-style-type: none">• Range: 1 to ± 1999.9 mV• Resolution: 0.1 mV (± 199.9 mV) / 1 mV (beyond) or better• Accuracy: ± 0.2 mV (± 199.9 mV) / 2 mV (beyond) or better• Calibration with at least two points (<u>specify how many</u>)• Temperature compensation: Automatic, 0.0 to 100.0 °C <p>Ions:</p> <ul style="list-style-type: none">• Capability to measure fluoride and other ions	1



Environmental Geochemistry Laboratory
Department of Civil Engineering
Indian Institute of Technology Kanpur

	<ul style="list-style-type: none">• Range: 0.01 ppm to 2000 ppm or wider• Resolution: 2 digits or better (user selectable)• Accuracy: 0.5 % Full scale (monovalent) / 1 % Full scale (divalent)• 4 or more calibration points <p>Temperature:</p> <ul style="list-style-type: none">• Range: 0.0 to 100.0 °C or higher• Resolution: 0.1 °C• Accuracy: $\pm 0.3^{\circ}\text{C}$ or better <p>Memory: Minimum 100 data sets Output: USB / IrDA Interface Adapter / RS232 / RS232C Power supply: 9 V DC adapter, long life compatible batteries, AC 230-240V 50/60 Hz Operating Temperature: 5-45 °C (ambient) Operating Conditions: 5-85 % (non-condensing)</p> <p>Weight: Preference for lower weight Diagnostic message display should be present.</p>	
2.	Compatible electrodes for parameters listed above (quote electrodes with model numbers and clearly specify whether these are included in the price of item 1):- <ul style="list-style-type: none">a) Stainless steel ATC probeb) Glass pH electrode, refillablec) Fluoride electroded) ORP electrode	2 sets
3	NIST traceable standards for calibrating above parameters. Quote separately their make, volumes (mL) and expiry dates, and clearly specify whether these are included in the price of item 1.	1 set

Terms and Conditions:

- The vendor should supply list of installation (minimum 5, in last two year) in India of the same model quoted against this enquiry preferably at IITK.
- Manufacturer should have appropriate certification.
- **If the Financial Bid is included in the Technical Bid, then the quotation will be rejected.**
- **Quotation should have minimum validity of 60 days from the date of opening.**
- Delivery period should be within 45 days from the receipt of the purchase order. Shorter delivery time may be given preference.



Environmental Geochemistry Laboratory
Department of Civil Engineering
Indian Institute of Technology Kanpur

- Taxes, packaging, forwarding freight charges, if any should be mentioned.
- Quotation should carry proprietary certificates and authorization letters/certificates.
- Prices should include installation and training of the equipment.
- Provide contact number/address for complaint, else quotes may be rejected.
- **The warranty period for the meter should be 3 years and at least 6 months for the electrodes from the date of installation.**
- The firms may also quote for optional accessories which will extend the capability or ease of use of the equipment.
- All quotations should be in the currency of the country of origin of the instrument and FOB and CIF, Delhi (if imported), and also converted to ₹.
- The Institute is exempted from excise duty and pays a nominal customs duty of ~5% under Govt. of India notifications 10/97 and 51196, respectively. Custom Duty exemption certificate under notification 51196 and road permit will be provided if applicable.
- Normal payment terms for the Institute will be applicable (90% on delivery of the items and the remaining 10% after satisfactory installation/ inspection).
- The Institute reserves the right for accepting and rejecting any quotations without assigning any reason thereof. **Also, the Institute reserves the right to reject or accept all or any of the offers made above.**

Thanking you,

Sincerely,

Dr. Abhas Singh

Assistant Professor, FB-306, Department of Civil Engineering,

I.I.T. Kanpur, Kanpur- 208016, Uttar Pradesh, India.