Sovan Lal Das, PhD

Assistant Professor Department of Mechanical Engineering Indian Institute of Technology Kanpur Kanpur – 208 016 (INDIA)



Phone : +91-512-259 7035 Fax : +91-512-2597408 E-mail : sovandas@iitk.ac.in

> IITK/ME/2013/SLD-02 Date: March 21, 2013

Sub: Purchase of Spectral Confocal Laser Scanning Microscope with provision of attaching and optical laser trap system

Kindly refer to the enquiry no. IITK/ME/2013/SLD-01 dated March 19, 2013. The terms and conditions of the tender should be read as per the following. Detailed technical specifications of the required system are available in the attached sheets. Please send sealed quotations for the same to:

Dr. Sovan Lal Das Department of Mechanical Engineering Indian Institute of Technology Kanpur Kanpur – 208016, India

Terms and Conditions:

- 1. The envelope must be inscribed with "Quotation for the Confocal Microscope System".
- 2. All quotations must reach by **March 26, 2013** at or before 1500hrs.
- 3. Quoted price should include the installation and training cost.
- 4. Maximum discount, if any, should be offered.
- 5. Quotations shall be submitted in two parts.
- i. Part-I (Technical): Should contain all technical details cum specifications of the offered solutions and be provided in a sealed envelope.
- ii. Part-II (Financial): Should contain the prices of the offered solutions along with any commercial terms and conditions and be provided in a sealed envelope. The prices should not be quoted in the technical bid.
- 6. Warranty and Support: Quotation should include warranty period for the instrument and the after sales service and application support capabilities.
- 7. FOB (indicating port of shipment) and CIF (New Delhi) values be separately mentioned if the item requires import.
- 8. The quote should cover insurance for transport up to IIT Kanpur.
- 9. Indian agency commission if applicable (should be certified by the principal if no agency commission is applicable) in case of import.
- 10. Authorization certificate from the principal if you are a local agent.
- 11. Terms and conditions for the payment, including the Banker's name of the principal and the account number, if any for electronic transfer.
- 12. Include proprietary item certificate if applicable.
- 13. Should provide technical literature to support the product.
- 14. Should provide users list with contact details in India.
- 15. The quotation must be valid for 90 days.
- 16. The delivery period should not be more than 8 weeks.
- 17. All prices are to be FOR IIT Kanpur.

With best regards,

Dr. Sovan Lal Das

<u>Specification for Spectral Confocal Laser Scanning Microscope with the provision for attaching Optical Laser Trap System</u>

The confocal microscope system should be of latest and modular technology suitable for biological science applications. System should be of high sensitivity detection capability to meet various challenging imaging needs of Biological samples including Live Cell imaging applications for FRAP, FRET, FLIP, photo activation and photo conversion experiments.

The system should be offered with the following configuration:

1) Fully Motorized & computer controlled Multiple port Inverted Fluorescence Research Microscope with a suitable laser port for Trapping experiments:

Bright field, Fluorescence and DIC observations with Motorized Z-focus drive with step resolution of 10 nm or better with dedicated TFT/LCD screen capable of controlling motorized functions of microscope. Motorized Universal Condenser with NA 0.55 or better, 6 position motorized FL filter wheel, 6 position motorized DIC nose piece. Programmable & Motorized XY Scanning Stage with universal sample holder for slides as well as 35 mm Petri dish for multiple location imaging. 12V /100w halogen illumination for BF & and 120W metal Halide Illuminator with high lifetime of 1500 hours for fluorescence. High resolution Confocal Grade objectives 20x/0.80 or better, 60/63x/1.40 oil. Band Pass fluorescent filters for DAPI, FITC/GFP & TRITC/Rhodamine. A suitable dichoric mirror for simultaneous optical trap/imaging should be incorporated in the motorized fluorescent turret. A high resolution color CCD Digital camera with 1.4 million pixel chip resolution, 2/3" CCD chip, FireWire IEEE 1394 connectivity, controlled by confocal software for high resolution fluorescence imaging.

2) Spectral Confocal Laser Scan head with built-in detectors:

High sensitivity confocal laser point scanning and detection unit with built-in spectral detectors for high efficient fluorescence signal collection. Capable of conventional intensity & Spectral based Confocal Imaging for complete visible range. Should be capable of simultaneous imaging of 2 fluorophores and at least 4 in sequential mode. Spectral dispersion of the emission light should be of latest technology with highly efficient spectral separation. Online separation and display of over-lapping emission signals through emission finger printing technique. Motorised & computer controlled continuously variable confocal pinhole with software control. High speed XY galvo scanner with 200 deg or better scan rotation with total scan flexibilities of Line, free hand curved line, XY, XYZ, XYZT and XYZTλ combinations. The laser scanner should have dual scan capability of fast scan for bleaching/photo activation & normal scan for Imaging, to conduct experiments like FRAP, FRET, PA/conversion. Scan resolution at least 2K x 2K for all channels and selected freely down to 4x1 pixels. Scan Zoom range 0.5X to 30X or more and adjustable in steps of 0.1. Scan speed of minimum 5 fps @ 512x512 pixel resolution and should increase to almost 150 fps at 512 X 16. Data acquisition and digitization capability with 8, 12 and 16 bit should be available.

3) Laser Modules:

The system should be offered with the following laser lines to excite the respective fluorochromes. Preference would be given to solid state lasers due to its long life and maintenance free operation.

- 488 nm for Alexa 488, FITC, GFP, Fluo 4, Cy2 fluorochromes.
- 555/559 nm for TRITC, Rhodamine, Texas Red, Cy3, PE, PI fluorochromes.
- 405 nm for DAPI, Hoechst fluorochromes.

All the lasers should be connected to the scan head through fiber optic cable. All the laser lines should be computer controlled for fast laser switching and attenuation in synchronization with the scanner.

4) Control Computer

Latest control computer with Core 2 Duo E8400 processor, RAM 4 GB DDR2-667,HDD: 500GB SATA II, DVD SuperMulti SATA +R/RW, SATI Fire GL V5600 512MB,Gigabit Ethernet, **Win Vista Ultimate 32/64bit**, USB 2.0, Firewire. Large 30" LCD TFT monitor or dual TFT monitors with 21".

5) System control and imaging software.

Software should be capable of controlling Motorized functions of microscope, scan head control, laser control, scanner control and Image acquisition & processing. Saving of all system parameters with the image for repeatable/reproducible imaging. Capability of Line, curved line, frame, Z-stack, Time series imaging. Photo-activation/conversion, FRAP & FRET imaging capabilities. Ion imaging with online ratio metric imaging and analysis. Standard geometry Measurements like length, areas, angles etc. including intensity measurements. 3D image reconstruction from a Z-stack image series. Colocalization and histogram analysis with individual parameters. Spectral un-mixing and emission fingerprinting technique should be standard feature of the software.

6) Installation and service support.

Bidders should clearly specify the after sales service and application support capabilities. Should provide all preinstallation requirements to have the system installed in ideal room conditions. Include the supply of appropriate Online UPS system to support the complete system including lasers. Provide a detailed list of users of the quoted system in India with contact details.

Dr. Sovan Lal Das