Phone: 0512 259 7128 Fax: (0512) 259 7395 E-mail: tarun@iitk.ac.in

Request for Quotation

Enquiry No: IITK/CESE/TG/2013-05 *Date: August 23, 2013*

Last Date: September 02, 2013

Sub: Call for quotation for "Laboratory Electric Oven"

Sealed quotations are invited for purchase of Laboratory Electric Oven.

Minimum Technical Specifications:

- 1. Temperature range upto 250°C with a minimum sensitivity of \pm 1°C.
- 2. It should have one path of circulating the hot air in between the inner chamber & insulation, forced air by the motorized blower (Perfect system) so that there is minimum temperature variation at any point in working space.
- 3. Temperature controlled by PID micro processor controller Dual Display with safety alarm.
- 4. Working chamber and outside body should be of good quality stainless steel (Full SS)
- 5. Volume: (Approx) 200 Litre or more.
- 6. Traceability calibration certificate of controller required

We are looking forward to receive your quotations, addressed to the undersigned. The deadline for submission of quotations is September, 02, 2013 at 4:00 pm.

Terms and Conditions:

- Quote should be made in two parts: Technical bid and financial bid separately in sealed envelopes.
- Financial bid for products whose technical bid is not acceptable will not be opened.
- ➤ If the financial bid is included in the technical bid then quotation will be summarily rejected.
- > The sealed envelopes with the quotes should be super scribed mentioning whether it is technical or financial bid.
- Taxes, packaging, forwarding freight charges, if any should be mentioned.
- Quotation should have a validity of a minimum of 60 days with complete technical details of the
- Warranty of at least 2 years is required.
- Maximum educational discount, if any, should be offered.
- All quotations should be in the currency of the country of origin of the instrument and CIF, Delhi.
- Quotation should carry proper certifications like Authorization, Proprietary Certificate etc.
- Institute is exempted for payment of Excise Duty under notification no. 10/97.
- The institute reserves the right for accepting and rejecting any quotations without assigning any reason thereof.

Thanking You

Date: August 23, 2013 Yours faithfully

Dr. Tarun Gupta

torus - Grusta

Associate Professor Department of Civil Engineering Indian Institute of Technology Kanpur Kanpur- 208016