

Indian Institute of Technology Kanpur

Department of Chemistry
Old Core Lab101C,
IIT Kanpur
Kanpur 208016
Uttar Pradesh, India



Raja Angamuthu
Assistant Professor
Tel: +91-512-259-6786
Fax: +91-512-259-7436
Email: Raja@iitk.ac.in

August 26, 2013

Enquiry No: IITK/CHM/RA/2013-14/002

Last Date: September 08, 2013

Quotations in sealed envelopes are invited

For the procurement of **ULTRASONIC BATH** comprising the following specifications, addressed to

Dr. Raja Angamuthu
Department of Chemistry
Indian Institute of Technology Kanpur
Kanpur 208016

Expected Technical Specifications	
Frequency	40 – 50 kHz
Heating	Yes
Digital Controls	Yes
Digital timer	Yes
Drain	No
Voltage Input	240V ($\pm 10\%$) at 50/60Hz
Tank Volume	2.5 – 3 Liters
Chemical resistant plastic body	Must
Cover	Yes

Price should include the transport until IIT Kanpur

Sincerely,

Raja Angamuthu