

INDIAN INSTITUTE OF TECHNOLOGY KANPUR
Department of Physics
Kanpur - 208016, India

Dr. Krishnacharya
Assistant Professor
Department of Physics
IIT Kanpur
Kanpur - 208016, India

kcharya@iitk.ac.in
Tel: +91-512-259-7968

Enquiry no.: PHY/KK/EQP/2013/1
Enquiry date: 08/01/2013
Closing date: 11/02/2013

Request for submission of quotation for “UV Ozone Cleaner”

Dear Sir/Madam,

Sealed quotation(s) are required on or before February 4th latest by 5 PM meeting all technical specifications as mentioned below:

1. Cleaning tray size (minimum): 25 cm × 25 cm × 5 cm
2. Intensity: > 30 mW/cm²
3. Digital control timer / display
4. Operating voltage: 220V

Terms and conditions:

Quotations should have a validity of a minimum of 60 days.

The equipment should be provided with a minimum warranty of 1 year.

Quotations are required in duplicate in a sealed envelope with enquiry number mentioned on the envelope.

The delivery period should be specifically stated.

The rate offered should be F.O.B (specify city).

Institute is exempted from payment of Excise Duty under notification no.10/97

Institute is entitled to avail concession rate of sales tax as admissible under Sub-sec 5 of

Sec 8 C.S.T Act 1956 applicable to Educational/Research institution in inter-state

Purchase.

Krishnacharya
Dept. Physics
IIT Kanpur