

**Enquiry letter for the Purchase of Laboratory Centrifuge**

Sealed quotations are invited from the manufacturer/authorized agents/resellers for the supply of Laboratory Centrifuge.

**Specifications:**

Capacity= 6×50 ml and 20×15 ml

Max. Speed RPM= 100-16000

Digital timer range= 0-59 min (Digital speed indicator)

Acceleration time= 0-30 second

Deceleration time= 0-30 second

Cooling= Non-refrigerated

Temperature range= 0 °C to +40 °C

Power Supply= 220-240 V, 50Hz single phase

**General Terms and conditions:**

- Quotes should have a minimum validity of 60 days
- The sealed envelopes with the quotes should be super-scribed with the enquiry number
- The delivery period should be specifically stated.
- Delivery on FOR Destination basis only.
- Product Warranty: 1 year

Sealed quotations subscribed with tender number must reach to us latest by May28, 2016 at the following address:

The Coordinator  
Center for Nanosciences  
Indian Institute of Technology Kanpur  
Kanpur 208016 INDIA