

**INDIAN INSTITUTE OF TECHNOLOGY KANPUR**  
**Department of Earth Sciences**

**\*Revised**

**Enquiry No: ES/RS/2015-2016/09**

**Date: 19.06.2015**

**Subject: Quotation for supply of components for a basic ratio sensor with the capacity to scan specific parts of the optical spectrum as mentioned below.**

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover in order to reach us by June 26, 2015 in the form of a hard copy to the address mentioned below. If you have any question please call Dr. Rajiv Sinha at 0512-2597317, email: [rsinha@iitk.ac.in](mailto:rsinha@iitk.ac.in)

**The prospective suppliers are required to send quotation in two parts in sealed envelopes, as "Technical Bid" and "Financial Bid".** The Technical Bid should contain detailed technical specification of the product being offered and should not mention any prices. The Financial Bid should include the detailed price quotation clearly including the cost of the equipment, taxes, service charges if any, shipping and handling charges. **The two separate and sealed envelopes should be clearly marked appropriately as "Technical Bid" and "Financial Bid". Kindly write the inquiry no on the top of envelop.**

**Terms and Conditions:-**

1. Maximum education discount, if any should be offered
2. Validity of quotation should be at least for 60 days
3. Prices should be on CIF and FOB separately (if imported)
4. Prices should include the installation and training cost
5. Normal payment terms for the Institute will be applicable (90% on delivery of the items and the remaining 10% after satisfactory installation/ inspection).
6. Quotation should carry proper certifications like agency certificate, proprietary certificate, etc.
7. Delivery should be made within 3 months.

**Technical specifications of Ratio Sensor:**

We require the components for a basic ratio sensor with the capacity to scan specific parts of the optical spectrum.

1. 2 small form factor monochrome cameras with max dimensions 60X60X60 mm
2. Recording bit-depth of 10-12 bits, 8 bit cameras not acceptable.
3. 18-20mm lenses
4. Global shutters
5. C mount for filters
6. Narrow-pass optical filters for 395, 450, 492, 520, 650, 800 and 852 nm
7. Control software

Dr. Rajiv Sinha  
Professor  
Department of Earth Sciences  
WLE Room no 202  
Phone: +91-512-2597317 (Office)  
Email: [rsinha@iitk.ac.in](mailto:rsinha@iitk.ac.in)