

Indian Institute of Technology - Kanpur
Department of Biological Sciences & Bioengineering

Enquiry Number: IITK/BSBE/SG/DBT/NC/2014-15/1, dated: 15-12-2014

Sub.: Inquiry for the supply of a laboratory freezer

Opening date: December 15, 2014 at 10:00 AM

Closing date: December 26, 2014 at 5:00 PM

Sealed quotes (technical bid and price bid separately sealed) are invited for the above mentioned laboratory product as per the specifications given in the next page.

Your quote should mention/include the following:

- Maximum discount if any should be offered and mentioned.
- Quoted price should include the cost for installation, warranty and required accessories
- Validity of the quote at least for 90 days.
- FOB (indicating port of shipment) and CIF (New Delhi) values should be quoted separately if import is required. For quotes in INR, the price quote should be for delivery at Kanpur.
- The quote should cover insurance for transport up to Kanpur.
- Indian agency commission if applicable (should be certified by the principal if no agency commission is applicable) in case of import.
- Authorization certificate from the principal if you are a local agent.
- Terms and conditions for the payment, including the banker's name of the principal and the account number, if any, for electronic transfer.
- Include proprietary item certificate if applicable.
- Technical literature to support your product (in technical bid).
- Users' list with contact address in technical bid.

Note: Only principal manufacturers or authorized representatives are requested to send the quote along with proper certificates. The envelope should be marked as "**Quote for laboratory freezer**".

Dr S.Ganesh
Department of Biological Sciences & Bioengineering
Indian Institute of Technology,
Kanpur 208016 (UP)

Laboratory Freezer [Quantity: 1]

Essential Technical Features:

Upright Deep Freezer with minimum five storage drawers
Should have a minimum capacity of 270 Liters
Should have Temp Range from -20°C to -30°C
Easy to use, slim & compact design
[maximum dimensions: 61.4cm W x 162cm H x 70.9cm D]
Microprocessor digital temperature display
Accurate temperature control by Microprocessor PID control system
Front-mounted display/control panel (with microprocessor)
Single electrical box for simpler servicing
Should have one access port
Should be mounted on 4 casters for movement
Environment-friendly refrigerant: CFC-free and HCFC-free
Should have pre-coated metal body
Door latch can accommodate a pad lock for added protection
Should have Memory backup
Should have safety features as Audio visual High/low temperature alarm and Power failure alarm
Should have Part replacement notification and Self diagnostics function
Should be an ISO 9001, 13485 & 14001 certified product

Warranty: Minimum one year warranty

Accessories:

Appropriate voltage stabilizer for the equipment should also be quoted

Preference will be given suppliers who have a proven track record of supplying such products to major research establishments in the country and good technical team for after sales service.