



**DEPARTMENT OF CHEMISTRY  
INDIAN INSTITUTE OF TECHNOLOGY  
KANPUR – 208 016  
INDIA**

**Tel:** 91-512-259 7251(O); 259 8261(R)

**Fax:** 91-512-259 7436/259 0260

**E-mail:** sprath@iitk.ac.in

**Dr. Sankar Prasad Rath**  
*Associate Professor*

---

**Date: 26-08-13**

**Enquiry No. IITK/CHM/SPR/13-14/02**

**Last date: 04 September 2013**

**Sub: Request for Quotation for Analytical Electronic Weighing Balance**

Please send a sealed quotation with all the technical details of the following item by September 4, 2013 to the following address.

Dr. S. P. Rath  
Department of Chemistry  
IIT Kanpur  
Kanpur-208016  
UP

**Detailed specifications:**

Readability : 0.01 mg

Repeatability : < 0.02 mg

Automatic calibration

Warranty: at least one year from the date of Installation

Quantity: 1